

**Pressure Systems, Inc.**  
Esterline Corporation  
34 Research Drive  
Hampton, VA 23666

Tel: 757-865-1243  
Fax: 757-766-2644  
[www.PressureSystems.com](http://www.PressureSystems.com)  
NYSE symbol: ESL

January 4, 2008

## **KPSI™ Transducers Cleaning Instructions and Certificate**

We are pleased you have selected our KPSI Transducers and we are confident that you've chosen the finest instruments in their class. Although these devices are virtually trouble-free, if it becomes necessary to return them to us for service, they must be cleaned and certified per the instructions on the back of this document.

Due to the potentially hazardous environments in which these products may operate, cleaning is mandatory so we can insure the safety of our service technicians. If cleaning is not performed and/or the certificate of compliance not returned with the product, we reserve the right to withhold service until appropriate cleaning, decontamination and certification have been completed. We will contact the sender for instructions regarding disposition of the equipment. Any costs for disposition will be borne by the sender.

Additional copies of cleaning instructions and Certificates of Compliance may be found on our website ([www.PressureSystems.com/cleaning.html](http://www.PressureSystems.com/cleaning.html)). For assistance or additional information, please call Customer Service at 800-328-3665. Thank you for your consideration.

Rick Hower, Quality Manager  
E-mail: [Rick.Hower@PressureSystems.com](mailto:Rick.Hower@PressureSystems.com)

## Cleaning Instructions

Before they can be serviced, transducers subjected to either biological or radioactive exposure or toxins, must be cleaned and disinfected. Biological contamination is presumed for any product that has been used with wastewater. Radioactive contamination is presumed for any product that has been used near any radioactive source.

When service is required, either at the user's facility or at PSI, the following steps must be taken to ensure the safety of our personnel:

1. Empty any containers or chambers, then thoroughly clean all matter from device.
2. In a manner appropriate to each device, decontaminate all exposed surfaces, including any containers. 70% isopropyl alcohol or a solution made fresh daily of ¼ cup bleach to 1 gallon tap water is suitable for most disinfection. Instruments used with wastewater may be disinfected with .5% Lysol if this is more convenient to the user.
3. The user shall take normal precautions to prevent radioactive contamination and must use appropriate decontamination procedures should exposure occur. If exposure has occurred, the customer must certify that decontamination has been accomplished and that no radioactivity is detectable by survey equipment.
4. Complete the Cleaning Certificate. If the product is being returned to Pressure Systems, pack it securely to prevent damage. **Please make a copy of the certificate and tape to outside of shipping box. The original certificate should be included with the product.** If the unit is to be serviced at the user's facility, keep the certificate with it.

Our personnel reserve the right to withhold service for any product until proper cleaning and decontamination have been accomplished and certified.

***Cleaning must be completed and certified on any product before returning it to PSI or before field service.***

-----

## CLEANING CERTIFICATE

Company \_\_\_\_\_ Department \_\_\_\_\_

Address \_\_\_\_\_

City \_\_\_\_\_ State \_\_\_\_\_ Zip \_\_\_\_\_

RMA # \_\_\_\_\_

Name/Model # \_\_\_\_\_

Serial Number \_\_\_\_\_

Cleaning certified by \_\_\_\_\_ Date \_\_\_\_\_

Potential Contaminants (if known) \_\_\_\_\_

Cleaning agent(s) used \_\_\_\_\_

If there has been radioactive exposure – is decontamination certified  Yes  No

Phone \_\_\_\_\_ Fax \_\_\_\_\_